

Riverside, Canal Road, Sowerby Bridge,  
West Yorkshire, HX6 2AY  
Tel. +44 (0)1422 829944  
Fax. +44 (0)1422 829945  
sales@contrec.co.uk



Flow Instrumentation

Company Reg. 7802714

VAT No. GB124 1591 39

## **EU DECLARATION OF CONFORMITY**

### **Conforming Apparatus**

**Rate Totaliser with Type Designation 202A**  
**Rate Totaliser with Type Designation 202D**  
**Batch Controller with Type Designations 214D**  
**Loop powered Indicators with Type Designation 220**  
**Process Indicator with Type Designation 250**

We hereby declare that, on the date the equipment accompanied by this declaration is placed on the market the equipment conforms with all technical and regulatory requirements of the listed directives below, when installed and operated according to the manual.

Directive 2014/34/EU – 20 April 2016



**Provision of the Directive fulfilled by the Equipment**

**II 2G Ex ia IIB T4 Gb**

**Notified Body for EC-Type Examination:**

Dekra EXAM GmbH – Notified body no. 0158  
Bochum Germany

**EC-type Examination Certificate:**

Dekra - BVS 15 ATEX E 106 X

**Notified Body for production:**

Sira – Notified body no. 0518  
Hawarden Industrial Park, Hawarden, Deeside. CH5 3US

**Harmonised Standards used:**

EN 60079-0:2012+A112013  
EN 60079-11:2012

**RoHS Directive: 2011/65/EU**

**EN 50581:2012** :- Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

**EMC Directive: 2014/30/EU**

**EN61326:2013** :- Electrical equipment for measurement, control and laboratory use – EMC requirements : Residential, Commercial & Light Industry Environment & Industrial Environment.

**Low Voltage Directive: 2014/35/EU**

**EN61010:2010** :- Safety requirements for electrical equipment for measurement, control and laboratory use.

T Casson, Production Director  
Contrec Ltd

Contrec are ISO 9001:2008 certified by Lloyds Certification

**contrec.co.uk**

**ACCURACY • QUALITY • PERFORMANCE**