

Riverside, Canal Road, Sowerby Bridge,
West Yorkshire, HX6 2AY
Tel. +44 (0)1422 829944
Fax. +44 (0)1422 829945
sales@contrec.co.uk



Flow Instrumentation

Company Reg. 7802714

VAT No. GB124 1591 39

EU DECLARATION OF CONFORMITY

Conforming Apparatus

Rate Totaliser with Type Designation 202A
Rate Totaliser with Type Designation 202D
Batch Controller with Type Designations 214D
Loop powered Indicators with Type Designation 220
Process Indicator with Type Designation 250

We hereby declare that, on the date the equipment accompanied by this declaration is placed on the market the equipment conforms with all technical and regulatory requirements of the listed directives below, when installed and operated according to the manual.

Directive 2014/34/EU – 20 April 2016



Provision of the Directive fulfilled by the Equipment

II 2G Ex ia IIB T4 Gb

Notified Body for EC-Type Examination:

Dekra EXAM GmbH – Notified body no. 0158
Bochum Germany

EC-type Examination Certificate:

Dekra - BVS 15 ATEX E 106 X

Notified Body for production:

CSA Group-Netherlands B.V. - Notified body no. 2813
Utrechtseweg 310 (B42) 6812AR ARNHEM

Harmonised Standards used:

EN 60079-0:2012+A11:2013
EN 60079-11:2012

RoHS Directive: 2011/65/EU

EN 50581:2012 :- Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

EMC Directive: 2014/30/EU

EN61326:2013 :- Electrical equipment for measurement, control and laboratory use – EMC requirements : Residential, Commercial & Light Industry Environment & Industrial Environment.

Low Voltage Directive: 2014/35/EU

EN61010:2010 :- Safety requirements for electrical equipment for measurement, control and laboratory use.

T Casson, Production Director
Contrec Ltd

Contrec are ISO 9001:2015 certified by
Lloyds Register

